

EU Quality Management System Certificate

Regulation (EU) 2017/745, Annex IX Chapter I and III

MDR 740894 R000

Manufacturer: tic Medizintechnik GmbH & Co. KG

Address:


Endelner Feld 9
Dorsten
46286
Germany

Single Registration Number: DE-MF-000033923

Scope: See attached **Device Schedule**

On the basis of our examination of the quality system in accordance with Regulation (EU) 2017/745, Annex IX Chapter I and III, the quality system meets the requirements of the Regulation. For the placing on the market of Class III devices, and Class IIb implantable devices that are not considered well-established technologies as specified in Article 52(4) an additional Annex IX Chapter II certificate is required.

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):



Graeme Tunbridge, Senior Vice President Global Regulatory & Quality

First Issue Date: **2022-01-24**

Current Issue Date: **2026-05-14**

Starting Validity Date: **2026-05-14**

Expiry Date: **2027-01-23**

...making excellence a habit.™

Page 1 of 3

Validity of this certificate is conditional on the Manufacturer's quality system being maintained to the requirements of the Regulation as demonstrated through the required surveillance activities of the Notified Body.
This certificate was issued electronically and is bound by the conditions of the contract.

EU Quality Management System Certificate


Regulation (EU) 2017/745, Annex IX Chapter I and III

MDR 740894 R000

Device Schedule: Class IIa, Custom-made and other devices

Device(s)	Risk Classification
Pessaries	Class IIa
SmartFlow	Class Im
UroPortNG	Class Im

For Class Im devices, the Notified Body conformity assessment is limited to the aspects relating to the conformity of the devices with the metrological requirements.



First Issue Date: **2022-01-24**

Current Issue Date: **2026-05-14**

Starting Validity Date: **2026-05-14**

Expiry Date: **2027-01-23**

...making excellence a habit.™

Page 2 of 3

Validity of this certificate is conditional on the Manufacturer's quality system being maintained to the requirements of the Regulation as demonstrated through the required surveillance activities of the Notified Body.
This certificate was issued electronically and is bound by the conditions of the contract.

EU Quality Management System Certificate

Regulation (EU) 2017/745, Annex IX Chapter I and III

MDR 740894 R000

Certificate History

(References to applicable Common Specifications, Harmonized Standards complied with, and the relevant test and audit reports that support any of the below certificate changes may be requested from Certificate.Verification@bsigroup.com)

Date	Reference Number	Action
2022-01-24	3338418	Issued
Current	30661117	Supplemented - New device for treatment genital descensus



First Issue Date: **2022-01-24**

Current Issue Date: **2026-05-14**

Starting Validity Date: **2026-05-14**

Expiry Date: **2027-01-23**

...making excellence a habit.™

Page 3 of 3

Validity of this certificate is conditional on the Manufacturer's quality system being maintained to the requirements of the Regulation as demonstrated through the required surveillance activities of the Notified Body.
This certificate was issued electronically and is bound by the conditions of the contract.